



SMART MEDIC

24 rue Feydeau
75002, Paris
FRANCE
18ret 88304048700018
TVA FR63883040487



**MASQUE MÉDICAL
TYPE IIR
NORME EN14683
CERTIFIÉ CE**

**NORME EN14683 : 2019
CERTIFIÉ CE**



SMART MEDIC

PHOTO





SMART MEDIC

PHOTO





SMART MEDIC

DESCRIPTIF

- **Masque médical non tissé(70%) et tissu fondu (30%)**
- **3 couches de protection**
- **Taille 17,5cm x 9,5cm**
- **Fixation par boucles élastiquées**
- **Efficacité de filtration bactérienne >98 %**
- **Couleur bleue**



SMART MEDIC

CERTIFICAT DE CONFORMITÉ

Nom du produit: Masque facial médical jetable (non stérile)
Spécification: élastique, Taille de Masque: 17.5x9.5cm
Norme exécutive: EN14683:2019
Contenu des ingrédients: 70% de tissu non tissé,
30% de tissu soufflé par fusion
Numéro de lot de production: 25004230
Date de production: 2020-10-14
Date d'expiration: 2022-10-13
Conditions de stockage: Stocker dans un
environnement ventilé et sec pour éviter l'humidité
et la moisissure.
Quantité: 50 pièces
Numéro d'employé de l'inspection de la qualité: QC01
Fabricant: GUANGDONG KINGFA SCI. & TECH. CO., LTD.
Adresse du fabricant: 28 DeLong Avenue, Shikok Town,
Qingcheng District, Qingyuan City,
Guangdong Province, China.

LE PRODUIT A RÉUSSI LE TEST,
POUR ÊTRE AUTORISÉ À QUITTER L'USINE

PRODUCT CERTIFICATE

合格证

Product Name: Disposable medical face mask (Non sterile)
Specification: Ear-loop 17.5x9.5cm
Executive standard: EN14683:2019
Content: 70% Non Woven Fabric, 30% Melt-Blown cloth
Production Lot No: 25004230
Production Date: 2020-10-14
Expiration Date: 2022-10-13
Storage conditions: store in a dry and ventilated
environment to avoid moisture and mildew
Number of products: 50 Pcs
Inspection No.: QC01
Manufacturers: GUANGDONG KINGFA SCI. & TECH. CO., LTD.
Manufacturer address: NO. 28, DeLong Avenue, Shijiao Town,
Qingcheng District, Qingyuan City, Guang
dong Province, China.

The Product passed the test, to be permitted to leave the factory

COVID-19

CONSIGNES D'UTILISATION

AVERTISSEMENT

Pour protéger votre santé et celle des autres, il est important de respecter ces consignes d'utilisation.

- Portez ce masque vous-même en contact avec d'autres personnes qui portent également votre masque.
- Vérifiez toujours que le masque est bien ajusté et couvre votre bouche et votre nez.
- Ce masque ne remplace pas les gestes barrières (lavage régulier des mains, distanciation physique, réduction des contacts avec d'autres personnes). Il ajoute une barrière physique, à utiliser notamment lorsque vous êtes en contact à bras armés avec d'autres personnes.

Comment mettre mon masque chirurgical? Comment mettre mon masque



Avant de le mettre :
01. Avant de toucher le masque, lavez-vous les mains avec de l'eau et du savon ou une solution hydro-alcoolique.
02. Inspectez le masque et assurez-vous qu'il n'y a pas de trous, de déchirure ou de gradations.
03. Il est recommandé de porter le masque sur une peau nue, en évitant le contact avec les cheveux.
04. Ne modifiez jamais le masque de quelque façon que ce soit.

Pour le mettre :
01. Tournez le masque dans la bonne direction (bord rigide en haut, face blanche vers vous).
02. Poussez-le sur votre visage.
03. Faites passer les boucles autour de vos oreilles (masques à boucles aux oreilles), ou attachez la partie sup. élastique (masques à lanières élastiques ou à attaches à nouer en haut et en bas).
04. Ajustez la bande pour le nez.
05. Attachez la partie inf. élastique si n'est pas (masques à lanières élastiques ou à attaches à nouer en haut et en bas).
06. Ajustez le masque de façon à recouvrir le nez, la bouche et que le bord inf. élastique recouvre votre menton.

Lorsque vous le portez :
01. Évitez de le toucher et de le déplacer.
02. Ne le mettez jamais en position d'attente sur le front ou sur le menton.

Il faut changer le masque :
01. Quand vous avez porté le masque 4 h.
02. Quand vous souhaitez boire ou manger.
03. Quand il devient difficile de respirer.
04. Si le masque est humide.
05. Si le masque est endommagé.
06. Si le masque est déformé et ne tient plus correctement contre votre visage.

Pour l'enlever :
01. Lavez-vous les mains avec de l'eau et du savon ou de la solution hydro-alcoolique.
02. D'échouez les lanières pour d'écoller le masque de votre visage ou d'enlever les noues, puis enlevez le masque en le maintenant par les attaches du haut.
03. Vous ne devez l'enlever qu'en touchant les bords, les attaches ou les boucles. Ne touchez pas la partie qui couvre votre bouche et votre nez.
04. Jetez-le tout de suite dans une poubelle qui se ferme.
05. Pour terminer, lavez-vous à nouveau les mains avec de l'eau et du savon ou de la solution hydro-alcoolique.



Ne pas réutiliser



Jeter le masque dans une poubelle fermée après usage (4h).
Put the masks in a closed bag after use (4h).

<h2 style="text-align: center;">EU Declaration of Conformity</h2> <p style="text-align: center;">according to the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL</p> <p style="text-align: center;"><i>Class I Medical Device</i> (non-sterile)</p>		
Manufacturer:	GUANGDONG KINGFA SCI.&TECH. CO., LTD.	
Address:	No.28 Delong Avenue, Shijiao Town, Qingcheng District, Qingyuan City, Guangdong Province, China	
Single Registration Number (SRN) of the Manufacturer:	--	
European Representative (ER):	Share Info GmbH	
Address:	Heerdter Lohweg 83, 40549 Düsseldorf	
Single Registration Number (SRN) of ER:	DE-AR-000005132	
We, the manufacturer, declare under our sole responsibility that		
the medical device(s)	Product Name:	Medical surgical mask
	Type/model, identification of product allowing traceability (Where applicable):	KF-C P01(R), KF-B P01(R), KF-K P01(R), KF-IP01(R), KF-B P01(R-BD), KF-L P01(R), KF-J P01(R), KF-B P08(R), KF-B P09(WK), KF-B P06(R), KF-B P06B(R)
	Intended Purpose:	Medical surgical masks providing barrier to minimize the direct transmission of infective agent between staff and patients are principally intended for use by healthcare professionals in an operating room and other medical settings with similar requirements, additionally, protect the wearer against splashes of potentially contaminated liquids. This is a single-use, non-sterile device.
	Classification: (Annex VIII of the MDR)	Class I Medical Device
	Basic UDI-DI:	KF-C P01(R): 697316340KF-CP01(R)83 KF-B P01(R): 697316340KF-BP01(R)7J KF-K P01(R): 697316340KF-KP01(R)CB KF-I P01(R): 697316340KF-IP01(R)B9 KF-B P01(R-BD): 697316340KF-BP01(R-BD)W4 KF-L P01(R): 697316340KF-LP01(R)CU KF-J P01(R): 697316340KF-JP01(R)BS KF-B P08(R): 697316340KF-BP08(R)93 KF-B P09(WK): 697316340KF-BP09(WK)TB KF-B P06(R): 697316340KF-BP06(R)8M KF-B P06B(R): 697316340KF-BP06B(R)R8
Conformity assessment route:		EU Declaration of Conformity + Technical Documentation (Annex II) + Technical Documentation on Post-Market Surveillance (Annex III)

is/are in conformity with the relevant provisions and requirements of the Council and the parliament regulation (EU) 2017/745 for medical device and all applicable harmonized standards and Common Specification. All supporting documents are retained under the premises of the manufacturer.

Applied harmonized standards and Common Specification	Regulation (EU) 2017/745	EN 14683: 2019+AC: 2019
	EN ISO 10993-1: 2018	EN 1041: 2008
	EN ISO 10993-5: 2009	EN ISO 14971: 2019
	EN ISO 10993-10: 2010	EN ISO 15223-1: 2016
	EN ISO 13485: 2016	EN 62366-1:2015
	MEDDEV 2.7.1: 2016	
Notified Body:	Not Applicable	
Address:	Not Applicable	
Identification Number:	Not Applicable	
EC Certificate(s):	Not Applicable	

Signed on:

Place: Qingyuan, China



2021-6-2

Signature (on behalf of the manufacturer) : GUANGDONG KINGFA SCI.&TECH CO., LTD.

Name of authorized signatory: Linanjin

Position held in the company: General Manager





中国认可
国际互认
检测
TESTING
CNAS L0599

Test Report

SL52045300144501TX

Date: October 26, 2020

Page 1 of 5

GUANGDONG KINGFA SCI.&TECH.CO.,LTD.
NO.28 DELONG AVENUE, SHIJIAO TOWN, QINGCHENG DISTRICT, QINGYUAN CITY, GUANGDONG
PROVINCE, CHINA

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A) Disposable medical mask/Medical surgical mask
Composition : (A) Polypropylene
Sample Color : (A) Blue
Style No. : KF-B P01(R)
Lot No. : 25004230
Manufacturer : GUANGDONG KINGFA SCI.&TECH.CO.,LTD.
Country of Destination : European Union

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Oct 13, 2020

Testing Period : Oct 13, 2020 - Oct 26, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo

Sara Guo (Account Executive)

Helen Xuan

Helen Xuan (Authorized Signatory)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-a-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/investigation report & certificate, please contact us at telephone: (86-21) 648871443, or email: CH.00000000@sgs.com

2nd Building No.888, Nanshan Road, Pudong District Shanghai, China 200235

T (86-21) 61402986

F (86-21) 64888783

www.sgs.com

中国·上海·浦东新区宜山路889号3号楼 邮编: 200235

T (86-21) 61402986

F (86-21) 64888783

e sgs.china@sgs.com

Member of the SGS Group (SGS SA)

Test Result

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

Clause 5.2 Performance Requirement

Clause 5.2.2 Bacterial Filtration Efficiency (BFE)

(EN 14683:2019+AC:2019 Annex B)

Sample: A
 Test Side : Inside
 Test Area : Approximately 60 cm²
 Flow Rate : 28.3 L/min
 Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.
 Dimensions of test specimen : ~171mm x 155mm
 Positive Control Average : 2582 CFU
 Negative Monitor Count : < 1 CFU
 Mean Particle Size : 3.0 ±0.3µm
 Test bacteria : Staphylococcus aureus ATCC 6538

Test Item	Specimen No.	Result
Bacterial Filtration Efficiency (BFE)	1	99.7%
	2	99.8%
	3	99.8%
	4	99.9%
	5	99.8%

Remark:

- 1) Performance Requirement: Type I ≥95%, Type II ≥98%, Type IIR ≥98%
- 2) The number of specimens that shall be tested is minimum 5, but can be greater and shall be increased if necessary to allow for an AQL(Acceptable Quality Level) of 4%.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic formal documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-a-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8387 1443, or email: Ch.Backcheck@sgs.com

2nd Building No.88, Yishan Road, Xuhui District Shanghai, China 200233 | (86-21) 61402666 | (86-21) 64656763 | www.sgs.com.cn
 中国·上海·徐汇区宜山路88号3号楼 邮编: 200233 | (86-21) 61402666 | (86-21) 64656763 | sgs.china@sgs.com

Clause 5.2.3 Breathability

(EN 14683 :2019+AC:2019 Annex C)

Sample: A

Test Side : Randomly test in different location (1 around and 4 away from the centric point) on each of the 5 masks

Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.

Test Area : 4.9 cm²

Flow Rate : 8 l/min

Specimen No.	Test Area No.	Different Pressure for each tested area (Pa/cm ²)	The average value for each test specimen (Pa/cm ²)
1	1-1	38.9	34
	1-2	38.3	
	1-3	29.4	
	1-4	32.4	
	1-5	32.2	
2	2-1	32.8	31
	2-2	31.4	
	2-3	30.6	
	2-4	29.6	
	2-5	32.6	
3	3-1	28.2	30
	3-2	29.2	
	3-3	32.0	
	3-4	30.7	
	3-5	27.6	
4	4-1	35.4	32
	4-2	34.1	
	4-3	33.2	
	4-4	30.0	
	4-5	29.2	
5	5-1	35.5	32
	5-2	35.0	
	5-3	31.2	
	5-4	28.6	
	5-5	29.9	

Remark:

- 1) Performance Requirement: Type I<40 Pa/cm², Type II<40 Pa/cm², Type IIR<60 Pa/cm²
- 2) The number of specimens that shall be tested is minimum 5, but can be greater and shall be increased if necessary to allow for an AQL(Acceptable Quality Level) of 4%.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-a-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-21) 64856763 or email: CN.Qc@sgs.com

1st Building No.888, Yinan Road, Pudong District, Shanghai, China 200233 t (86-21) 61403666 f (86-21) 64856763 www.sgs.com.cn
中国·上海·浦东新区宜山路888号3号楼 邮编: 200233 t (86-21) 61403666 f (86-21) 64856763 e: sgs.china@sgs.com

Member of the SGS Group (SGS SA)

Clause 5.2.4 Splash Resistance
(ISO 22609 :2004)

Sample: A

Test Blood Pressure

: 16.0kPa

Pre-Conditioning

: Minimum of 4 hours at 21±5°C and 85±5% R.H.

Distance of the mask to the tip of cannula

: 300±10mm

Test Specimen#	Penetration on inside surface	Conclusion	Test Specimen#	Penetration on inside surface	Conclusion
1	None Seen	Pass	17	None Seen	Pass
2	None Seen	Pass	18	None Seen	Pass
3	None Seen	Pass	19	None Seen	Pass
4	None Seen	Pass	20	None Seen	Pass
5	None Seen	Pass	21	None Seen	Pass
6	None Seen	Pass	22	None Seen	Pass
7	None Seen	Pass	23	None Seen	Pass
8	None Seen	Pass	24	None Seen	Pass
9	None Seen	Pass	25	None Seen	Pass
10	None Seen	Pass	26	None Seen	Pass
11	None Seen	Pass	27	None Seen	Pass
12	None Seen	Pass	28	None Seen	Pass
13	None Seen	Pass	29	None Seen	Pass
14	None Seen	Pass	30	None Seen	Pass
15	None Seen	Pass	31	None Seen	Pass
16	None Seen	Pass	32	None Seen	Pass
Number of Pass:			32		
Overall result:			Acceptable		

Remark:

- 1) Performance Requirement Type I: N/A, Type II: N/A, Type IIR: ≥16.0kPa
- 2) Test was conducted within 60s after removal from conditioning chamber.
- 3) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/sgs/terms-and-conditions.sgs> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/terms-and-conditions/terms-e-document.sgs>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8397 1443, or email: CN.DocsCheck@sgs.com

2nd Building, No. 388, Nishan Road, Pudong District, Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64866763 www.sgschina.com.cn
中国 - 上海 - 浦东新区宜山路388号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64866763 e sgs.china@sgs.com

Member of the SGS Group (SGS SA)

Sample Photo



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.sgs> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-758) 83871443, or email: CN_Decehack@sgs.com

J Building No.888, Yishan Road, Pudong District, Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64966763 www.sgs.com.cn
中国·上海·徐汇区宜山路888号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64966763 e: sgs.china@sgs.com

Member of the SGS Group (SGS SA)